ACTIV SCREW.





CANNULATED Ø6.5 MM SCREWS

ACTIV SCREW

Indication for use: The implants of the Activ Fuse range are intended for bone reconstruction of the ankle joint in adults including fractures fixation and arthrodeses of the ankle, distal tibia, talus, and calcaneus.

Contraindications:

- Pregnancy.
- Acute or chronic, local or systemic infections.
- Allergy to one of the materials used or sensitivity to foreign bodies.

A COMPREHENSIVE RANGE OF SCREWS AND INSTRUMENTS

→ SCREW FEATURES



→ INSTRUMENT FEATURES

- A comprehensive and specific kit for Ø6.5 mm cannulated screws (see page 4 for more information).
- A protection sleeve (ANC754) for soft tissues preservation and the whole surgery guided through a Ø2.5 mm L230 mm pin (33.0225.230).
- Two types of handles (ANC442 and ANC443) for an ergonomic use.
- A cannulated Jacobs adaptor (ANC1585) for large couplings of drill bit (ANC727) and 2 in 1 instrument (ANC730).



SURGICAL TECHNIQUE

EXAMPLE OF APPLICATION: TIBIO-TALAR ARTHRODESIS



.....

- 1. Perform the desired incision and preparation.
- 2. Through the incision, insert the soft tissue protection sleeve (ANC754) up to the bone.

The tissue protection sleeve can be held with the handle (ANC442).

NB: the whole procedure can be done through the tissue protector.

3. Screw the Ø2.5 mm pin guide (ANC735) into the tissue protection sleeve and insert a Ø2.5 mm pin (33.0225.230).

Then, remove the pin guide from the tissue protector.



4. Guide the length gauge (ANC733) along the pin and through the tissue protection sleeve up to the bone surface. Read the insertion depth on the length gauge (at the end of the pin).



5. Perform a soft preparation of the cortex to bury the head of the screw into the bone with the countersink part of the 2 in 1 instrument (ANC730).



- 6. Drill through the tissue protection sleeve with the Ø4.6 mm cannulated drill bit (ANC727).
- NB: the drilling depth can be checked at the rear of the tissue protector using the markings on the drill bit.



7. Insert the desired screw through the tissue protector with the screwdriver part of the 2 in 1 instrument (ANC730).



- 8. Repeat the whole procedure for the other screws.
- NB: a washer (WASH-T6.5-ST) can be used in the case of poor bone quality.



Tissue protector is not compatible with a washer.

Final tightening of the screws must be performed by hand.





REFERENCES

→ SCREWS & WASHER

IMPLANTS	
Ref.	Description
H2.6QT6.5LxxD-ST*	Ø6.5 mm compressive screw - cannula Ø2.6 - short thread - L40 to 120 mm - STERILE (5 mm increments)
H2.6JT6.5LxxD-ST*	Ø6.5 mm compressive screw - cannula Ø2.6 - long thread - L45 to 120 mm - STERILE (5 mm increments)
H2.6CT6.5LxxD-ST*	Ø6.5 mm standard cortical screw - cannula Ø2.6 - full thread - L40 to 120 mm - STERILE (5 mm increments)
WASH-T6.5-ST	Compression washer for Ø6.5 mm compressive screws - STERILE



*Remark : please note that the «xx» in the references represents the length of the screw.

Ex: the reference for the Ø6.5 mm standard cortical screw - cannula Ø2.6 - full thread - L40 mm - STERILE is «H2.6CT6.5L40D-ST»

→ INSTRUMENTS & KIT DESCRIPTION


INSTRUMENTS			
Ref.	Description	Qty	
ANC442	Ø11 mm cannulated quick coupling straight handle	1	
ANC443	Ø11 mm cannulated quick coupling T-handle	1	
ANC727	Ø4.6 mm drill bit - cannula Ø2.6 mm - L260 mm	1	
ANC730	2 in 1 : T25 screwdriver Ø2.6 mm cannulated - Ø8.0 mm countersink	2	
ANC733	Length gauge for pin Ø2.5 mm - L210 mm	1	
ANC735	Ø2.5 mm pin guide	1	
ANC754	Soft tissues protection sleeve	1	
ANC1585	Jacobs cannulated quick coupling adaptor for Ø11 mm instruments	1	
33.0225.230	Pin Ø2.5 - L230 mm	6	

REMOVAL INSTRUMENTS

If you have to remove ACTIV SCREW implants, make sure to order the **Newclip Technics** removal instruments:

> - ANC442: Ø11 mm cannulated quick coupling straight handle - ANC730: 2 in 1: T25 screwdriver Ø2.6 mm cannulated - Ø8.0 mm countersink

Non contractual pictures.



ANC1143/B Activ Screw - Ø6.5 mm cannulated instrumentation set - Base

This information is intended to demonstrate the Newclip Technics portfolio of medical devices. Always refer to the package insert, product label and/or user instructions including cleaning and sterilization before using any Newclip Technics product. These products must be handled and/or implanted by trained and qualified staff who have read the instructions before use. A surgeon must always rely on her or his own professional clinical judgement when deciding whether to use a particular product when treating a particular patient. Product availability is subject to the regulatory or medical practices that govern individual markets. Please contact your Newclip Technics representative if you have questions about the availability of Newclip Technics products in your area.